

State of California
AIR RESOURCES BOARD

EXECUTIVE ORDER I-10-028

Relating to Independent Contractor Approval under
California Code of Regulations, Title 17, Section 91207

Datalab

WHEREAS, the Air Resources Board (ARB), pursuant to California Health and Safety Code section 41512, has established the procedures contained in California Code of Regulations, title 17, sections 91200-91220, to allow the use of independent testers for compliance tests required by ARB;

WHEREAS, it has been determined that Datalab meets the requirements of ARB for conducting ARB Test Methods 1, and 4 pursuant to California Code of Regulations, title 17, sections 91200-91220, when the following conditions are met:

1. Datalab calibrates its temperature gauges in accordance with section 4.3 of ARB Test Method 2, and establishes and maintains a log of the calibrations;
2. Datalab calibrates its dry gas meters as required by section 5.3 of ARB Test Method 5, and establishes and maintains a log of the calibrations;
3. Datalab measures barometric pressure as described in section 2.5 of ARB Test Method 2. It must calibrate its barometer in accordance with section 4.4 of ARB Test Method 2, and establishes and maintains a log of the calibrations; and

WHEREAS, ARB's Executive Officer, pursuant to Health and Safety Code section 39516, issued Executive Order G-02-008, delegating to the Chief of ARB's Monitoring and Laboratory Division the authority to approve independent testers in accordance with California Code of Regulations, title 17, sections 91200-91220.

NOW, THEREFORE, I, Alberto Ayala, Chief of ARB's Monitoring and Laboratory Division, order that Datalab is granted an approval, from the date of execution of this order, until June 30, 2011, to conduct the tests listed above, subject to compliance with California Code of Regulations, title 17, sections 91200-91220.

BE IT FURTHER ORDERED that during the approved period the Executive Officer or his authorized representative may field audit one or more tests conducted pursuant to this order for each type of testing listed above.

Executed at Sacramento, California this 27th day of July 2010.

Original signed by

Alberto Ayala, Ph.D., M.S.E.
Chief, Monitoring and Laboratory Division